

TECHNICAL EVALUATION REPORT (SUMMARY)

1. Name of Procuring Agency: University of Health Sciences Lahore
 2. Method of Procurement: Single Stage two Envelopes
 3. Title of Procurement: Procurement of Laboratory Equipment for Microbiology Department, University of Health Sciences, Lahore
 4. Tender Inquiry No: UHS/Pur/1395/21
 5. Date & Time of Bid Closing: 22-03-2022 at 11:00 AM
 6. Date & Time of Bid Opening: 22-03-2022 at 11:30 AM
 7. No. of bids received: 06

| Sr. No | Item | Specifications | Bidders | | | | | Worldwide Scientific |
|--------|----------------------------|--|---------------------|--|------------------|-----------------|----------------|----------------------|
| | | | New Chemical Centre | Muslim Medical Services | S.U. Enterprises | SMS Diagnostics | Emerging Tech. | |
| 1 | Ultra Low Freezer -70°C | Type Ultra Low Upright Origin(Manufacturing) USA/UK/EU Capacity 19-20 cu .ft. (500-550L) Temperature Range -50°C to -80°C Temperature Accuracy °C ±1 Temperature Uniformity °C ±5 Holding capacity 2 in. Boxes (400) Rack capacity 8-16 Insulation Vacuum Insulation Panels (VIP) + CFC free Polyurethane Foaming Material (Interior) Stainless Steel (Painted) Alarms High/low temperature, power failure, door open, sensor failure, | Qualified | Not Qualified (Chinese Origin) | Qualified | Qualified | Qualified | Qualified |


| | | | | | | | |
|--|---|--|--|--|--|--|--|
| | <p>high ambient temperature, hot condenser and low battery alarms</p> <p>Controller Type Microprocessor Control, LED Display, USB Data Export</p> <p>Refrigerant R290 (1st stage), R170 + R290 mix (2nd stage)</p> <p>Compressor Double compressors: Rotary Type (High stage) Hemetic type (Low stage)</p> <p>Defrost: Manual</p> <p>Safety Features: Automatic vent port</p> <p>CE Certified</p> | | | | | | |
|--|---|--|--|--|--|--|--|

All technical bid was evaluated by EndUsers against the laid specifications and comparative statement (summary/detail) is being submitted to Technical committee for further necessary action.



Head of Microbiology Department


No Complaint Shall be entertained after 07 days of display of Technical Evaluation report.



Dr. Uruj Zehra
Convener



Lt.Col. (R). Dr. Khalid Rahim
Member



Dr. Allah Rakha
Member