


## TECHNICAL EVALUATION REPORT (SUMMARY)

1. Name of Procuring Agency: University of Health Sciences Lahore  
 2. Method of Procurement: Single Stage two Envelopes  
 3. Title of Procurement: Purchase of Laboratory Research Equipment for Morbid Anatomy & Histopathology Department, University of Health Sciences, Lahore  
 4. Tender Inquiry No: UHS/Pur/1384/2021  
 5. Date & Time of Bid Closing: 21-03-2022 at 11:00 AM  
 6. Date & Time of Bid Opening: 21-03-2022 at 11:30 AM  
 7. Nos. of Bids Received: 07

Sr #	Item	BOQs	Emerging	Info Tech	SU Enter	Sindh MS	WWS	Science Tech	Science Tech	New Chem
1	<b>ULTRA LOW FREEZE R -86°C with 5 KVA Stabilizer</b>	Temperature range should be -40°to -86°C, ambient temperature -16 - 32 °C, Cooling Performance: -86°C, Cooling Method: Direct Cooling, Controller: Microprocessor, Capacity should be 700L, until should attain -80 °C in less than 03 hours, Dual door, lock with key, Touch screen with high / low temp alarm, system sensor error, system failure, main board communication error and display can be operated with gloves. Interior finishing stainless steel, Shelves Number should be 6, level number 6, shelf material should be stainless steel, total number of racks 12, digital plus, USB port, data logger functions. Electric power 1290 W and Voltage/Frequency 230/50, Temperature display screen with lock system, and 5 KVA stabilizer.	Quoted as per BOQs  Qualified	Quoted as per BOQs  Qualified	Quoted as per BOQs  Qualified	Vender has quoted 578 Liter capacity which is not as per required specification therefore,  <b>NOT QUALIFIED</b>	Quoted as per BOQs  Qualified	Quoted as per BOQs  Qualified	Quoted as per BOQs  Qualified	Quoted as per BOQs  Qualified

Air Conditioner Inverter (02 Tons)	Cooling Rated Capacity :6400 (2800-7000) W Cooling Rated Power Consumption : 2020 (730-2600) W Cooling Running Current :8.8 (3.4-11.4) A Seer Cooling :3.6 Heating Rated Capacity :6400 (2900-7200) W Heating Rated Power Consumption :2020 (800-2700) W Heating Running Current :8.8 (3.6-11.8) A Low Voltage Operation :Yes	Not Quoted	Quoted as per BOQs Qualified	Not Quoted	Not Quoted	Not Quoted	Not Quoted	Not Quoted	Quoted as per BOQs Qualified
---------------------------------------------	----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	---------------	------------------------------------	------------	------------	---------------	------------	---------------	------------------------------------

All technical bid was evaluated by EndUsers against the laid specifications and comparative statement (summary/detail) is being submitted to Technical committee for further necessary action.



**Prof. Dr. Nadia Naseem**

HOD Morbid Anatomy & Histopathology


**No Complaint Shall be entertained after 07 days of display of Technical Evaluation report.**



**Dr. Uruj Zehra**  
Convener



**Lt.Col. ( R ). Dr. Khalid Rahim**  
Member



**Dr. Allah Rakha**  
Member